

9 May 2019

To: Recipients of M100, 29th ed.
From: Jennifer K. Adams, MT(ASCP), MSHA
Vice President, Standards and Quality
Subject: Correction

This notice is intended to inform users of a correction made to CLSI document M100, *Performance Standards for Antimicrobial Susceptibility Testing*, 29th ed. The correction is described below and shown as highlighted text in the table excerpt.

Table 2C. Zone Diameter and MIC Breakpoints for *Staphylococcus* spp.:

In the Testing Conditions box, in the "Medium" section, the special media conditions for testing daptomycin were omitted. The text has been corrected to include "CAMHB supplemented to 50 µg/mL calcium for daptomycin."

Table 2C. Zone Diameter and MIC Breakpoints for *Staphylococcus* spp.

Testing Conditions	
Medium:	Disk diffusion: MHA Broth dilution: CAMHB; CAMHB + 2% NaCl for oxacillin; CAMHB supplemented to 50 µg/mL calcium for daptomycin Agar dilution: MHA; MHA + 2% NaCl for oxacillin. NOTE: Agar dilution has not been validated for daptomycin.
Inoculum:	Colony suspension, equivalent to a 0.5 McFarland standard
Incubation:	35°C ± 2°C; ambient air Disk diffusion: 16–18 hours; 24 hours (for ceftioxin when testing <i>Staphylococcus</i> spp., excluding <i>S. aureus</i> , <i>S. lugdunensis</i> , <i>S. pseudintermedius</i> , and <i>S. schleiferi</i>) Dilution methods: 16–20 hours; 24 hours for oxacillin and vancomycin Testing at temperatures above 35°C may not detect MRS.

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service (customerservice@clsi.org).

We appreciate your commitment to CLSI and regret any inconvenience.